



CERTIFICATE OF MANUFACTURE/ANALYSIS

Product Size:	<u>3.78 Liters (1 Gallon)</u>	Product Name:	<u>Glass Klenz</u>
Manufacture Date:	<u>March 23, 2017</u>	Product ID:	<u>1114-08</u>
Exp. Date:	<u>N/A</u>	Lot No.:	<u>286597</u>

I. The product listed above is made per the current issue of the following documentation:

Raw Material Quality Control Specifications
Finished Product Master Formula
Manufacturing Instructions
Packaging Specifications
Finished Product Quality Control Specifications

Copies of this documentation are on file at STERIS Corporation, St. Louis Operations and will be provided as required to the applicable regulatory agencies.

II. The product listed above has been analyzed. The release specifications and results are shown below.

<u>Test Name</u>	<u>Method No.</u>	<u>Results</u>	<u>Specification</u>
Color	110	Orange	Orange
Clarity	107	Clear	Clear
pH (1% w/w soln)	106	11.3	10.9 – 11.6
Total Alkalinity (as Na ₂ O)	4	4.1%	3.8 – 4.3%
Total Chelant as EDTA	220	19.5%	18.8 – 20.6%
Specific Gravity (@ 25°C)	114	1.136	1.123 – 1.146

Based on these results the product is released for shipment to all STERIS Corporation distribution points.

Signed:

Date: March 28, 2017

Quality Representative