

Certificate Of Processing

Prepared for MICRONOVA MFG INC

STERIS

Gamma Process Run ID 116868A

Product Code

Product Lot Number

Quantity

UOM

VWBL-40TRIR

014784-000

60

CS

Other Information:

0002704-00

Processing Run Start Date/Time: 01-Aug-2019 04:45:05 pm

Approx. Downtime (hours): 0.12

Processing Run End Date/Time: 01-Aug-2019 11:56:12 pm

Minimum Specified Dose (kGy): 40.0

Minimum Delivered Dose (kGy): 43.7

Maximum Specified Dose (kGy): 60.0

Maximum Delivered Dose (kGy): 53.8

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Signature Manifest

Reviewed and E-Signed By

Joanna Jimenez (QS & RC Technician)

Document Content Revision: 1

Signed On 8/2/2019 at 9:15 AM

UTC / GMT Offset (hh:mm): -7:00

Processing Location:

STERIS

1000 South Sarah Place

Ontario, CA 91761

Phone: 909-390-9942

Fax: 909-390-7854

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485, and in alignment with EN ANSI/AAMI/ISO 11137. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.