



CERTIFICATE OF CONFORMANCE AND STERILIZATION

Date: 15-Nov-19

This certification is provided as full assurance that the following product code was manufactured in accordance with prescribed procedures and specifications.

Product/Material Number:	C2-1212IR
Product Description:	STERILESORB* C/P 12INx12IN
Lot Number:	172818
Irradiation Run Number:	2549014
Irradiation Dose Delivered:	32.1kGy to 40.6kGy
Manufacture Date:	10/2019
Expiration Date:	10/2021

The above product has been Gamma Irradiated to Sterility Assurance Level of 10^{-6} in accordance with the guidelines of the Association for the Advancement of Medical Instrumentation.

A handwritten signature in black ink, appearing to read "C. Inthavongsa".

Champagne Inthavongsa
Quality Supervisor
cinthavongsa@contecinc.com

R55480102

11/11/2019 15:25:21 GMT
Page 1 of 1

Customer Name:	FG Clean Wipes- Chicopee, MA.	Processing Facility:	Charlotte Gamma	Work Order #	2549014
P.O.#	38104			Sales Order #	2361983
25-50 kGy		LYM2, Gamma Treatment	Irradiation Date/Time:	11/08/2019 12:17:00 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	1040	CA	C2-1212IR	Wiping Cloth	172818	NA
	1040	CA	Total			

Processing Summary

Op#	Operation Name	Location	Date/Time In (GMT)	Date/Time Out (GMT)
400.00	Receive Product	UNPROC	11/01/2019 14:36:00	11/06/2019 21:45:41
410.00	Prepare Product	UNPROC	11/06/2019 21:45:41	11/07/2019 22:17:51
420.00	Load Product - A123	UNPROC	11/07/2019 22:17:51	11/08/2019 12:17:00
430.00	Process Product	UNPROC	11/08/2019 12:17:00	11/08/2019 12:42:18
440.00	Unload Product	UNPROC	11/08/2019 12:42:18	11/08/2019 22:40:25
450.00	Collect Run Data	UNPROC	11/08/2019 22:40:25	11/09/2019 08:30:51

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	50.0 kGy	32.1 KGY	Pass	FGDIAZ	11/09/2019 08:29:45 GMT
	Reason Code Test					FAUSTINO G. DIAZ	
450.00	Maximum Dose	25.0 kGy	50.0 kGy	40.6 KGY	Pass	FGDIAZ	11/09/2019 08:30:01 GMT
	Reason Code Test					FAUSTINO G. DIAZ	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: WENDY KOBLISH
Reason: Work Order Completions

Date: 11/11/2019 15:23:19 GMT