

Certificate of Conformity

VWR Catalogue Number	See List
Description	VWR®, Test tubes without rim, thick walled
Country of Origin	Manufactured in India
Date of Issue (yyyy-mm-dd)	2023-10-16

Catalogue Number	Length [mm]	Ø ext [mm]	Wall Thickness [mm]
212-0734	40	8	0.8 – 1.0
212-0735	70	10	0.8 – 1.0
212-0736	75	12	0.8 – 1.0
212-0737	100	10	0.8 – 1.0
212-0738	100	12	0.8 – 1.0
212-0739	100	16	0.8 – 1.0
212-0740	100	14	0.8 – 1.0
212-0741	130	14	0.8 – 1.0
212-0742	160	16	0.8 – 1.0
212-0743	180	18	0.8 – 1.0

Quality System Compliance

Products are manufactured under the **ISO 9001:2015** standard and calibrated in **ISO 17025:2017** certified Laboratory. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

The product is in conformance with the requirements of the following standards and regulations:

Technical standards for the products: NA

Calibration procedure: NA

Material	The products are made from Borosilicate 7.0 expansion.
Sterilization	Not Applicable.
ATP Assay	Not Applicable.
DNase & RNase Free	Not Applicable.
BSE/TSE	No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE).
Cytotoxicity	Not Applicable.
Non-Pyrogenic Statement	Not Applicable.
Latex Statement	Not Applicable.
BPA Statement	Not Applicable.
DEHP Statement	Not Applicable.
RoHS	Not Applicable.
REACH Statement	Not Applicable.
Storage Conditions	Room temperature.
Shelf Life	No predefined shelf life.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.

Version History

Sl.no	Version number	Description	Month & Year
1	1.0	Created General CoQ using new layout	03-2023