



TERUMO Europe N.V.
Interleuvenlaan 40
3001 Leuven
Belgium
Tel: +32 16 38 12 11

Leuven, 16 December 2024

CONFORMITY CERTIFICATE

PRODUCT DESCRIPTION	LOTN°	MANUF. DATE	EXPIRY DATE
SR*FF2419	220107A	2022/01/07	2026/12

Tests performed :

- sterility control	conform
- research of pyrogens (pyrogen free)	conform

Tests performed are conform with Japanese Pharmacopoeia.

K. Verhaert
Quality Responsible

THIS CERTIFICATE IS ISSUED BY TERUMO EUROPE IN THE NAME OF TERUMO CORPORATION JAPAN