



CERTIFICATE OF CONFORMANCE

Product Part Number: MZCRMS-10IR
Lot Number: 047474-000
Irradiation Reference Number: 0010021-00
Lot Quantity: 134 CASE(S)
Date of manufacture: 01/03/2025*
Shelf Life: 60 MONTHS
Description: MEGASORB MOP, BLUE, IRR'D (12/CS)
Size -
Width: 6.5"
Length: 15"
Height: N/A

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

3/31/2025

Authorized Signature
Quality Control Department

Date

CC: With Product
Customer File
C of C File

***Refer to Certificate of Processing for irradiation complete date**

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980



A Sotera Health company

Certificate of Processing

STERIGENICS 3125 Wichita Ct. Fort Worth TX 76140
TEL 817 293-0999 FAX 817 293-2933 www.sterigenics.com

RSS480102

01/20/2025 18:39:42 GMT
Page 1 of 1

Customer Name: **Micronova Manufacturing**
P.O.# **10021**

Processing Facility: **Fort Worth**

Work Order# **3690303**
Sales Order# **3720716**

25-45 kGy

**MICRO225/45, Gamma
Treatment**

Irradiation Date/Time: **01/17/2025 00:33:00 GMT**

Irradiation Cell: **CELLS**

SO Line#	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	134	CA	MZCRMS-10IR 109_11011	MEGASORB MOP.MICROFIBER.I'RRD HI: TAS LO: 0.SCI	047474-000	9
	Dose Map					

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	Signed By..... User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.9 KGY	Pass	DJOHNSON1	01/19/2025 04:44:47 GMT
	Reason Code Test					DAVID JOHNSON	
450.00	Maximum Dose	25.0kGy	45.0 kGy	37.9 KGY	Pass	DJOHNSON1	01/19/2025 04:45:06 GMT
	Reason Code Test					DAVID JOHNSON	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: **SUFYAN IBRAHIM**
Reason: **Work Order Completions**

Date: **01/20/2025 18:39:01 GMT**

APPROVED

By Jenicah Baquir at 8:23 am, Mar 31, 2025