



CERTIFICATE OF CONFORMANCE

Product Part Number: CSLB-12WHIR
Lot Number: 022829-000
Irradiation Reference Number: 0004625-00
Lot Quantity: 5 CASE(S)
Date of manufacture: 02/03/2021*
Shelf Life: 24 MONTHS
Description: WHITE MICROSEAL TAPE, IRR'D (72/CS)
Size -
Width: ½"
Length: 18 YARDS
Height: N/A

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

A handwritten signature in blue ink that reads "Sabito".

02/09/2021

Authorized Signature
Quality Control Department

Date

CC: With Product
Customer File
C of C File

***Refer to Certificate of Processing for irradiation complete date**

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

STERIS: Gamma Certificate Of Processing

Prepared for: MICRONOVA MFG INC (2727)

Gamma Process Run ID: 1113-1870A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
CSLB-12WHIR	022829-000	5	Case

Processing Run Start Date: 08-Feb-21 01:10 PM Approx. Downtime (hours): 0.48

Processing Run End Date: 08-Feb-21 09:15 PM

Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	26.8
Maximum Specified Dose (kGy):	40.0	Maximum Delivered Dose (kGy):	35.4

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Other Information

0004625-00

Signature Manifest

Reviewed and E-Signed By: **Joanna Jimenez (QS & RC Technician)**

Date/Time E-Signed: 2021-02-09 11:51 AM

Document Content Revision: 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location

1000 Sarah Place
Ontario CA 91761
United States