
HACH LANGE GmbH

Willstätterstr. 11
40549 Düsseldorf
Germany
Tel. + 49 (0) 211 5288-0
Fax. + 49 (0) 211 5288-210
info-de@hach.com
www.hach.com

DECLARATION OF CONFORMITY

I, the undersigned : **Bernd SEIDL**
HACH LANGE GmbH
Königsweg 10
D-14163 Berlin
Germany

Position in the company : **Deputy Manager Calibration Lab.**

Declare that the products : **Code** 2283549.99
Designation pH 7.00
Lot 23321
Expiration date 05/31/2025

Comply with our specifications when sold:

Buffer solution : pH 7.00 ± 0.02 (25°C)

Our pH measurements are traceable to CRM (Certified Reference Material) from NIST (National Institute of Standards and Technology) and/or others national standards

Berlin, November 29, 2023



Produced by:

HACH LANGE GmbH
Königsweg 10
14163 Berlin
Germany
www.hach.com