



Certificate of Conformance and Sterility

This certification is provided as full assurance that the following product code and lot number were manufactured in accordance via a Quality System certified to ISO 9001 with prescribed procedures and specifications.

Product Code:	C2-99IR/25
Lot Number:	193120
Product Description:	SterileSorb™, Sterile hydroentangled polyester/cellulose wipes, 9" x 9" (23 x 23cm)
Manufacturing Date:	07/2023
Expiration Date:	07/2025

Sterility

Gamma Processing Run ID:	3385009
Irradiation Dose Specified:	25.0 kGy to 50.0 kGy
Irradiation Dose Delivered:	33.6 kGy to 43.3 kGy
Sterility Assurance Level:	10 ⁻⁶ (SAL determined by test method AAMI/ISO 11137)

A handwritten signature in black ink, appearing to read "C. Duffee".

QC Supervisor
coc@contecinc.com

Wednesday, August 02, 2023

NORTH AMERICA

Contec, Inc.
525 Locust Grove
Spartanburg, SC 29303 USA
tel: +1 864 503 8333
fax: +1 864 503 8444

EUROPE

Contec Europe
ZI Du Prat, Avenue Paul Duplex
56000 VANNES France
tel: +33 (0)2 97 43 76 98
fax: +33 (0) 97 43 76 86

ASIA

Contec Cleanroom Technology (Suzhou) Co., Ltd.
17 Longyun Road
Suzhou Industrial Park, Suzhou 215024 China
tel: +86 512 6274 4050
fax: +86 512 6274 4051

web: www.contecinc.com
e-mail: info@contecinc.com

Page 1 of 1

DID: 14921D



Certificate of Processing

STERIGENICS 10811 Withers Cove Park Dr. Charlotte NC 28278
TEL 704 588-6877 FAX 704 588-8355 www.sterigenics.com

R55480102

07/31/2023 13:49:27 GMT
Page 1 of 1

Customer Name:	FG Clean Wipes- Chicopee, MA.	Processing Facility:	Charlotte Gamma	Work Order #	3385009
P.O.#	P043049			Sales Order #	3355786
25-50 kGy		LYM2, Gamma Treatment	Irradiation Date/Time:	07/29/2023 11:07:00 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	560	CA	C2-99JR/25	Wiping Cloth	193120	N/A
102.000	480	CA	C2-1212IR	Wiping Cloth	193117	N/A
	1040	CA	Total			

Processing Summary

Op#	Operation Name	Location	Date/Time In (GMT)	Date/Time Out (GMT)
400.00	Receive Product	UNPROC	07/27/2023 14:53:00	07/27/2023 16:28:52
410.00	Prepare Product	UNPROC	07/27/2023 16:28:52	07/27/2023 18:23:23
420.00	Load Product - A123	UNPROC	07/27/2023 18:23:23	07/29/2023 11:07:00
430.00	Process Product	UNPROC	07/29/2023 11:07:00	07/29/2023 11:36:25
440.00	Unload Product	UNPROC	07/29/2023 11:36:25	07/29/2023 22:55:23
450.00	Collect Run Data	UNPROC	07/29/2023 22:55:23	07/29/2023 23:31:20

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By ----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	50.0 kGy	33.6 KGY	Pass	TDIAZ	07/29/2023 23:30:18 GMT
	Reason Code Test					Faustino (Tino) Díaz	
450.00	Maximum Dose	25.0 kGy	50.0 kGy	43.3 KGY	Pass	TDIAZ	07/29/2023 23:31:10 GMT
	Reason Code Test					Faustino (Tino) Díaz	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: BRITTANY LONG
Reason: Work Order Completions

Date: 07/31/2023 13:48:26 GMT