

Certificate Of Processing

Prepared for MICRONOVA MFG INC

STERIS

Gamma Process Run ID 99870A

Product Code

BKL1-1818IR

Product Lot Number

Z-2321

Quantity

128

UOM

CS

Other Information:

p.o# P-5554

Processing Run Start Date/Time: 09-Mar-2017 10:30:36 pm

Approx. Downtime (hours): 0.00

Processing Run End Date/Time: 10-Mar-2017 02:53:39 am

Minimum Specified Dose (kGy): 40.0

Minimum Delivered Dose (kGy): 43.5

Maximum Specified Dose (kGy): 60.0

Maximum Delivered Dose (kGy): 57.8

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Signature Manifest

Reviewed and E-Signed By

Adrian Martinez (QS & RC Technician)

Document Content Revision: 1

Signed On 3/10/2017 at 8:01 AM

UTC / GMT Offset (hh:mm): -8:00

Processing Location:

STERIS Isomedix Services

1000 South Sarah Place

Ontario, CA 91761

Phone: 909-390-9942

Fax: 909-390-7854

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485, and in alignment with EN ANSI/AAMI/ISO 11137. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.