

CODAN FRANCE SARL
49 RUE DE ROHRWILLER,
BP 40073
67242 BISCHWILLER CEDEX
REPUBLIC OF FRANCE

TEST CERTIFICATE

Product : 1 ML CODAN U100 LUER. NEUTRAL GAS-PAPER WITH
FULL IN-LINE PRINT. INSERT IN BOX.

Quantity....: 12.800

CODAN REF: 62.1640

LOT: J86479-1

Invoice No.....: 447332

Steril.Date: 2024-08

Results

DS/EN ISO 11138

Biological Indicators for Ethylene Oxide Sterilization **Passed**

European Pharmacopoeia

2.6.14. Bacterial Endotoxins
(except oral tip one)

Passed

**The validation of the sterilization process
was performed according to:**

DS/EN ISO 11135-1

Sterilization of Health Care Products - Ethylene Oxide -
Requirements for Development, Validation and Routine Control
of a Sterilization Process for Medical Devices.

EN 556

Sterilization of Medical Devices -
Requirements for Medical Devices to be labelled "Sterile".

This certificate has been issued
electronically and is valid
without signature or stamp.

Date : Aug.21,24