



Certificate

PREDNISONE TABLETS

USP Catalog No.: 1559505

USP Lot No.: R132B0

(10 mg nominal prednisone content per tablet)
FOR DISSOLUTION PERFORMANCE VERIFICATION TEST (PVT)

Period of validity: This certificate of USP Prednisone Tablets Lot R132B0 is valid through 31 December 2022.

The USP Prednisone Tablets RS is provided for use in the *Performance Verification Test* for USP Apparatus 1 and 2 with 1-Liter vessels in the USP General Test Chapter on DISSOLUTION <711> and DRUG RELEASE <724>, APPARATUS SUITABILITY. Store in a dry place. Store the tablets at controlled room temperature not exceeding 25°.

Dissolution Medium: We recommend preparing the medium as follows:

Heat a suitable amount of water, while stirring gently to about 41-45°. Filter under vacuum through a 0.45-μm-porosity filter into a suitable filtering flask equipped with a stirring device. Seal the flask and continue to apply vacuum while stirring for an additional five minutes. Measured vacuum should be less than 100 mbar. The temperature of the *Dissolution medium* should not fall below 37° prior to the initiation of the test.

Procedure [See DISSOLUTION <711> in the current USP]: Determine the quantity of prednisone, C₂₁H₂₆O₅, dissolved at 30 minutes, in each vessel, expressed as percent of the labeled amount. Use 499 g of *Dissolution Medium* (which corresponds to 500 mL), where possible the medium should not be stirred prior to the initiation of the test for the purpose of equilibration, and conduct the test at 37°. Operate each apparatus at 50-rpm speed. Withdraw an aliquot of sample solution at 30 minutes and filter immediately. Measure the amount of prednisone dissolved from filtered portions of the sample aliquots at 242 nm in comparison with a solution of known concentration of USP Prednisone Reference Standard.

Notes: An amount of alcohol not to exceed 5% of the total volume of the standard solution may be used to bring the prednisone reference standard into solution. The filtering method must not cause adsorptive loss of drug. Bias introduced by automated methods is to be avoided. If equipment is dedicated for use with only one apparatus (basket or paddle), then performance verification is only required for that apparatus. At the time of use, peel back the paper-backed lidding to remove the tablets from the blister card.

Test Interpretation: Laboratory can choose either of the test schemes listed below.

Single-Stage Test

The following are step-by-step instructions for the Single-Stage test.

1. For each position in the assembly, test one USP Prednisone Tablets RS, and record the percent dissolved at the sampling time point specified. Transform the percent dissolved results to the natural log scale, determine the mean and variance. For assemblies with 12 or 14 positions (12 or 14 dissolution vessels), no further testing is required.
2. For assemblies with fewer than 12 positions, repeat Step 1 with an additional set of tablets. Again after transforming the percent dissolved results to the natural log scale, determine the mean and variance.
3. Calculate the average of the two means and of the two variances obtained in Steps 1 and 2. (Use the results from Step 1 alone for assemblies that have 12 or 14 positions.)
4. Convert the results of Step 3 to a geometric mean (GM) and percent coefficient of variation (%CV). See calculation example below for more detail.

Copyright 2018 The United States Pharmacopeial Convention. All rights reserved.

USP Certificate

USP Template No.: CERT1-07

Template Effective Date: August 31, 2021

Page 1 of 5



5. Compare the results of Step 4 to the **Single-Stage** acceptance ranges in Table 1. The GM must not fall outside the limits, and the %CV must not be greater than the limit. If both meet the criteria, the assembly has passed the PVT.

Optional Two-Stage Test

A laboratory may choose to implement the PVT as a Two-Stage test in case of assemblies with less than 12 positions. The Two-Stage test is a statistically valid means of allowing the possibility of stopping the test at the first stage with a penalty. The following are step-by-step instructions for the two-stage test.

1. For each position in the assembly, test one USP Prednisone Tablets RS, and record the percent dissolved at the sampling time point specified. After transforming the percent dissolved results to the natural log scale, determine the mean and variance.
2. Convert the results of Step 1 to a GM and %CV, and compare to the **1st Stage of Two Stages** acceptance ranges in Table 1. The GM must not fall outside the limits, and the %CV must not be greater than the limit. For calculation of the GM and %CV, see calculation example for more detail.
3. If results of Step 2 satisfy both acceptance criteria, the assembly has passed the PVT. Otherwise continue to Step 4. (see *note 1*).
4. Repeat Step 1 with an additional set of tablets and after transforming the percent dissolved results to the natural log scale determine the mean and variance for the data obtained at this step.
5. Average the two means and two variances obtained in Steps 1 and 4.
6. Convert the results of Step 5 to a geometric mean (GM) and percent coefficient of variation (%CV). For calculation of the GM and %CV, see calculation example for more detail.
7. Compare the results of Step 6 to the **2nd Stage of Two Stages** acceptance ranges in Table 1. The GM must not fall outside the limits, and the %CV must not be greater than the limit. If both meet the acceptance criteria, the assembly has passed the PVT.

In order to comply with the requirements of ASTM E29, all limit values in Table 1 are expressed with two significant figures.

Table 1: Performance Verification Test (PVT) limits (values apply only to Lot R132B0)

Apparatus	# of vessels	Single-Stage		Two-Stage			
				1 st Stage of Two Stages		2 nd Stage of Two Stages	
		GM*	%CV	GM*	%CV	GM*	%CV
1	6	47-76	16	51-70	12	47-76	15
	7		15	52-70			
	8		15				
	12	47-77	16	Not Applicable			
	14	47-76	15				
2	6	27-37	8.3	28-35	6.2	27-37	7.7
	7		8.1				7.5
	8		7.9				7.4
	12		8.2	Not Applicable			
	14		8.0				

* Percent of the labeled amount of prednisone dissolved at 30 minutes at 50 rpm



Calculation example (expressed as Microsoft Excel® worksheet functions):

Run 1: $x_1, x_2, \dots, x_n$ in natural log scale: $\text{Ln } x_1, \text{Ln } x_2, \dots, \text{Ln } x_n$

Run 2: $x_{n+1}, x_{n+2}, \dots, x_{2n}$ in natural log scale: $\text{Ln } x_{n+1}, \text{Ln } x_{n+2}, \dots, \text{Ln } x_{2n}$

1st Stage of Two-Stage for $n=6, 7, 8$ and Single-Stage for $n=12, 14$:

$$\text{GM1} = \exp(\text{average}(\text{Ln } x_1 : \text{Ln } x_n))$$

$$\%CV1 = 100 * \sqrt{\exp(\text{var}(\text{Ln } x_1 : \text{Ln } x_n)) - 1}$$

Single-Stage or 2nd Stage of Two-Stage for $n=6, 7, 8$:

$$\text{GM} = \exp(\text{average}((\text{average}(\text{Ln } x_1 : \text{Ln } x_n)), (\text{average}(\text{Ln } x_{n+1} : \text{Ln } x_{2n})))) = \exp(\text{average}(\text{Ln } x_1 : \text{Ln } x_{2n}))$$

$$\%CV = 100 * \sqrt{\exp(\text{average}((\text{var}(\text{Ln } x_1 : \text{Ln } x_n)), (\text{var}(\text{Ln } x_{n+1} : \text{Ln } x_{2n})))) - 1}$$

exp: exponential (e^x)

var: variance

sqrt: square root

*: multiply

100: conversion factor to percentage

Note 1:

There are circumstances when the %CV after the first stage equals or exceeds the value in the **Futility Factor** table (without rounding), then it is impossible to meet the %CV criterion after the second stage. The lab can stop after the first stage (run). However, after any adjustments to equipment, test procedure, and so on, the PVT must be restarted with a new first run.

Futility Factor

(%CV at or above value given, second stage testing will not produce passing result)

Apparatus	No. of Vessels		
	6	7	8
1	21	21	21
2	11	11	10

**LABEL TEXT****REFERENCE STANDARD****PREDNISONE TABLETS 30 Tablets**

The nominal weight of prednisone in each tablet is 10 mg. Do not push tablets through foil backing. To remove tablets from blister, peel foil. Use only whole tablets. Store in a dry place. Store at controlled room temperature not exceeding 25°C. Unused or unopened blister strips should be kept in the secondary package.



Danger! Causes eye irritation. Suspected of damaging fertility or the unborn child. Causes damage to organs (endocrine system) through prolonged or repeated exposure.

Obtain special instructions before use. Do not handle until all safety precautions have been read and understood. Do not breathe dust. Wash thoroughly after handling. Wear protective gloves/protective clothing/eye protection/face protection. If in eyes: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. If eye irritation persists: Get medical advice/attention. If exposed or concerned: Get medical advice/attention. Get medical advice/attention if you feel unwell. Store locked up. Dispose of contents/container in accordance with local/regional/national/international regulations.

See certificate for any additional information.

LOT: R132B0

USP, 12601 Twinbrook Pkwy, Rockville, MD, +1-301-881-0666
Cat. No. 1559505 Material mfd. in Spain



For use with specified USP compendial tests. Not for use as a drug.
See SDS prior to use at www.usp.org/sds.

Quality Assurance

Certificate Version History

Version Number	Date	Reasons for Change
00	30-JUL-2021	First issue
01	18-MAR-2022	The Valid Use Date (VUD) has been extended from 30 June 2022 to 30 September 2022.
02	26-JUL-2022	The Valid Use Date (VUD) has been extended from 30 September 2022 to 31 December 2022.

**Assigned Value**

Please refer to the USP Reference Standard label and/or USP Certificate for the assigned value of the specific lot. If an assigned value is not included on the label or Certificate, the lot was developed for qualitative USP compendial use and an assigned value will not be provided.

Valid Use Date

It is the responsibility of the user to ascertain that a particular lot of a USP Reference Standard has official status either as a "Current Lot" or as a "Previous Lot" within the assigned valid use date. The online USP Reference Standards Catalog and the online USP Store at www.usp.org are updated daily. USP recommends referring to one of these sources prior to using a USP RS to make sure the lot is valid for use.

Storage

Storage conditions are lot-specific and may change from one lot to another. The storage condition for an unopened USP Reference Standard is provided on the container label only, not on the Safety Data Sheet. If no specific directions or limitations are on the label, conditions of storage include storage at room temperature and protection from moisture, light, freezing, and excessive heat. See General Chapter <659> in the USP-NF Online for storage and handling definitions.

Instructions for Use

Follow the instructions provided on this Certificate, on the label of the USP Reference Standard, and in the associated USP documentary standard(s). Please refer to General Chapter <11> for additional information.

Non-USP Compendial Use

USP Reference Standards are intended only for use in analytical or laboratory applications generally as specified in USP compendia. They are not for use in humans or animals as drugs or medical devices. It may be possible to use a USP RS outside of its associated USP compendial applications; however, it is the responsibility of the user to determine the suitability of the USP RS for a non-USP use.

LEGAL NOTICE

USP WARRANTS GOOD TITLE TO USP REFERENCE STANDARDS ON DISPATCH FROM USP. THE FOREGOING WARRANTY IS IN LIEU OF ANY OTHER WARRANTIES, EXPRESS OR IMPLIED, INCLUDING WITHOUT LIMITATION ANY WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE, OR ANY WARRANTY THAT THE PRODUCTS, INCLUDING THIS CERTIFICATE, ARE OF MERCHANTABLE QUALITY. USP'S LIABILITY ARISING OUT OF OR RELATING TO THE SUPPLY OF USP REFERENCE STANDARDS AND THIS CERTIFICATE SHALL IN NO EVENT INCLUDE LOSS OF PROFITS, COST OF PROCURING SUBSTITUTE GOODS OR SERVICES, OR ANY INCIDENTAL, INDIRECT, OR CONSEQUENTIAL DAMAGES OF ANY KIND, EVEN IF USP IS AWARE OF THE POSSIBILITY OF SUCH DAMAGES. WITHOUT LIMITING THE GENERALITY OF THE FOREGOING, USP DOES NOT WARRANT THAT THE USE OR RESALE OF USP REFERENCE STANDARDS, INCLUDING THEIR USE TO PERFORM TESTS AND ASSAYS PUBLISHED BY USP, WILL NOT INFRINGE UNITED STATES OR ANY OTHER PATENTS.

USP Reference Standards are not intended for use as drugs, dietary supplements, or as medical devices.

This certificate may not be reproduced without the express written permission of USP.



Dissolution Performance Verification Standard– **Prednisone**



September 30, 2022

New standard announcement!

In our continued effort towards providing a comprehensive performance qualification of dissolution instruments, we are launching a new product for Performance Verification Testing (PVT) of apparatus 1 and 2.

The new product, Dissolution Performance Verification Standard - Prednisone (Item No. 1222818), is targeted to launch in November of 2022 and will officially replace the current Reference Standard, USP Prednisone Tablets (Item No. 1559505), on May 1, 2023.

For questions regarding this Notice of Intent to Revise, please email Margareth Marques, Senior Principal Scientist, mrm@usp.org.

For questions on the upcoming DVPS - Prednisone standard, please email RSTech@usp.org.

[Notice of Intent to Revise](#)



[FAQs](#)

