

Certificate of Quality

VWR Catalogue Number	See List
Description	VWR® Dissecting Scissors, Sharp/Blunt Tip
Country of Origin	Manufactured in Pakistan
Date of Issue (yyyy-mm-dd)	2025-11-11

VWR Catalogue Number	Length	Unit
82027-588	140 mm (5½")	Each
		Pack of 10
		Case of 50
82027-594	165 mm (6½")	Each
		Pack of 10
		Case of 50

Quality System Compliance

This certificate confirms that the product has been sourced from a supplier or manufacturer who has declared their Quality Management System (QMS) complies with the [ISO 9001:2015 & ISO13485:2016 & F-899](#) standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable product specifications.

Product Specifications

Material	AISI 410.
Sterilization	Not applicable.
ATP Assay	Not applicable.
DNase & RNase Free	Not applicable.

BSE/TSE	No use of any raw material produced from, or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE).
Cytotoxicity	Not applicable.
Non-Pyrogenic Statement	Not applicable.
Latex Statement	The products are free from Latex.
BPA Statement	Not applicable.
DEHP Statement	Not applicable.
RoHS	Not applicable.
REACH Statement	Not applicable.
Storage Conditions	Store in dry place, avoid direct sunlight.
Shelf Life	5 years from date of manufacture.

Disclaimer: VWR'S Quality Management System (QMS) is ISO 9001:2015 certified as a wholesale distributor of finished products, ensuring effective oversight of finished product handling, traceability and distribution processes. VWR is not a manufacturer and does not engage in manufacturing, co- manufacturing, or processing activities such as design, packaging, labelling, testing, or quality release. Consequently, the supplier or manufacturer's QMS certification may differ from VWR's.

VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. This product is intended for research use only. It is not for use in diagnostic procedures, therapeutic applications, or any human or clinical use. This document has been produced electronically and is valid without a signature.

Version History

Sl.no	Version number	Description	Month & Year
1	1.0	Created General CoQ using new layout	November 2025