


Test Report

– Eppendorf Biopur –

Safe-Lock Tubes 2.0 mL



100 EPPENDORF TUBES®
Eppendorf BIOPUR®
centrifugation up to 22 000 x g

LOT J193073R

2025-12

STERILE R

Order no.: 0030 121.597
Eppendorf AG
22331 Hamburg · Germany
Email: eppendorf@eppendorf.com

• **Human DNA**

Method: *real-time* PCR acc. to SOP 99880/V2
Limit: < 2 pg; less than one human cell

• **Bacterial DNA**

Method: *real-time* PCR acc. to SOP 100567/V1
Limit: < 50 fg; less than 10 *E. coli* cells

• **DNase**

Method: DNA Digestion acc. to SOP 100570/V1
Limit: Not detectable
(LOD: 1.0×10^{-6} Kunitz units)

• **RNase**

Method: RNA Digestion acc. to SOP 100571/V1
Limit: Not detectable
(LOD: 1.0×10^{-9} Kunitz units)

• **PCR inhibition**

Method: *real-time* PCR acc. to SOP 100559/V1
Limit: Less than 10 targets amplifiable

• **Sterility**

Method: Sterilization with SAL 10^{-6} by
☒ Irradiation according DIN EN ISO 11137:2015* or ☐ Ethylene oxide*
*Please see attached label for the standard method of sterilization for this product

Test method: USP, Ph. Eur. 8.0 / 2.6.12

• **Endotoxin**

Method: Kinetic-turbidimetric LAL test (Ph. Eur. 8.0 / 2.6.14)
Limit: < 0.001 EU/mL

• **ATP**

Method: Pre-sterilization test / Bioburden test (Ph. Eur. 8.0 / 2.6.12)
Limit: < 5.5×10^{-12} mg

Customer: Eppendorf AG, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 23.11.20

Testing of the above described lot showed conformity within the limits of detection.

01.12.20

Date of release

Epp 2011 7754

Registration number



Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.