

**TERUMO (PHILIPPINES) CORPORATION**

124 East Main Ave., Laguna Technopark, Biñan, Laguna, Philippines
Tel. No. (049) 541-2111 • Fax No. (049) 541-2121

19QC-C-2809
February 10, 2020

CERTIFICATE OF CONFORMITY

PRODUCT CODE	LOT NO.	STERILIZATION DATE	EXPIRY DATE
SR+OX2225C1	200202SC	2020-02	2025-01-31

- I. Visual Inspection : Conform
- II. Functional Inspection : Conform
- III. Sterilization : Conform
- IV. Bacterial Endotoxin Test : Conform

The visual and functional inspections are checked in accordance with Terumo specifications based on ISO standards.

Sterilization method : Electron Beam Sterilization

Sterilization dose is determined and audited periodically according to ISO11137.

Sterilization Assurance Level (SAL) is 10^{-6} .

The bacterial endotoxin limit for medical device is tested in accordance to USP standard.

TERUMO (PHILIPPINES) CORPORATION has established and applied a quality system for manufacturing of medical devices.

Report No.: 12031274 006

Registration No.: SX 60136578 0001

Furnish requirement: EN ISO 13485:2016

Valid until: 2022-02-11

Maricris Arcilla
Quality Control Manager