

Certificate of Conformity

VWR Microscope Slide, With Cut Edges, Frosted

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: 631-1554

Lot No: 081817

Date of Manufacture: 18th Aug 2017

Length (in mm)	Width (in mm)	Pack
76	26	50

Quality System Compliance: Products are manufactured under the ISO 9001:2008 standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

Material: Float Glass

Thickness: 1.00 to 1.20 mm

Refractive index: 1.53 ± 0.02 of refractive index of light at the green mercury e-line
(wavelength = 547.07nm) (ISO 8037-2)

Wettability: The central parts of the two surfaces of the slides should be wetted equably and slightly dehydration of the edges is allowed. (ISO 8037-2)

Certificate of Conformity

BSE/TSE Statement: The above product contains no ingredients of animal origin and no material derived from, or exposed to animals affected by or under quarantine for Transmitting, Animal Spongiform Encephalopathy (TSE)/ Bovine Spongiform Encephalopathy (BSE)

This document has been produced electronically and is valid without a signature

Date of issue: **April 2022**