

Certificate of Conformity Finntip™ Sterile

Product Code: 9401113
Product Name: Finntip 1000, sterile
Lot Number: 22013F0
Exp: 2027-01
Manuf. Date 2022-01

We hereby confirm that these high quality Finntip™ pipette tips are manufactured under controlled environment and are certified by validation of the manufacturing process to be free of the following contaminants:

Human DNA	< 1.9×10^{-11} g/tip
DNase	< 9.4×10^{-4} U/tip
RNase	< 6.5×10^{-8} U/tip
Endotoxin	< 0.05 EU/ml

The sterility (SAL 10⁻⁶) is ensured by irradiating this product according to ISO 11137 guidelines.

In Joensuu, 12th October 2023

For



Elina Maaranen, Quality Manager
Laboratory Plastics Essentials
Thermo Fisher Scientific Oy