

Certificate of Conformity

VWR Catalogue Number	631-0130
Description	Rectangular cover glasses, #1, 18 mm length, 24 mm width
Lot Number	24317
Country of Origin	Manufactured in China
Date of Issue (yyyy-mm-dd)	2026-03-27

Quality System Compliance

This certificate confirms that the product has been sourced from a supplier or manufacturer who has declared their Quality Management System (QMS) complies with [ISO 13485:2016](#).

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

Material	White hydrolytic class 1 glass.
Sterilization	Not Applicable.
ATP Assay	Not Applicable.
DNase & RNase Free	This product is free of any detectable DNase/RNase contamination.
BSE/TSE	Not Applicable.
Cytotoxicity	Not Applicable.
Non-Pyrogenic Statement	Product is non-pyrogenic.
Latex Statement	This product is not made from natural rubber latex.
BPA Statement	Not Applicable.
DEHP Statement	Not Applicable.

RoHS	No substances (Lead, Cadmium, Mercury, Hexavalent Chromium (Cr6+), Poly Brominated Biphenyls (PBB), Poly Brominated Diphenyl ethers (PBDE), Bis(2-Ethyhexyl) Phthalate (DEHP), Benzyl Butyl Phthalate (BBP), Dibutyl Phthalate (DBP), Diisobutyl Phthalate (DIBP)) are used in manufacturing the raw materials and final product. No routinely analyse is performed.
REACH Statement	No substances as listed on >Substance of Very High Concern< candidate list are used in manufacturing of the raw materials used to produce this product. No routinely analyse is performed.
Storage Conditions	Store under dry and clean condition at ambient temperature (15–30° C).
Shelf Life	Not Applicable.

Disclaimer: VWR's Quality Management System (QMS) is ISO 9001:2015 certified as a wholesale distributor of finished products, ensuring effective oversight of finished product handling, traceability, and distribution processes. VWR is not a manufacture and does not engage in manufacturing, co-manufacturing or processing activities such as design, packaging, labeling, testing or quality release. Consequently, the supplier or manufacturer's QMS certification may differ from VWR's.

VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. This product is intended for research use only. It is not for use in diagnostic procedures, therapeutic applications, or any human or clinical use.

This document has been produced electronically and is valid without a signature.