

BRAND GMBH + CO KG | Otto-Schott-Str. 25 | 97877 Wertheim | Germany

BRAND GMBH + CO KG
P.O. Box
97861 Wertheim
Germany
T +49 9342 808 0
info@brand.de
www.brand.de

Wertheim, 16.12.2024

Certificate of Quality

Product: Membrane filter pore size 3.0 µm

Cat. No.: 26056

Lot No.: 24H/230661103

Our membrane filters are manufactured in accordance with the Current Good Manufacturing Practice Standards and in a facility whose Quality Management System is certified for compliance with the DIN EN ISO 9001 Quality System Standard and DIN EN ISO 13485.

During manufacture each unit is tested 100% for membrane and housing integrity by a leakage test under automatic conditions. Before packing each unit has been checked by visual inspection.

The product is biosafe according to the USP Class VI Biological Tests for Plastics.

Following 100% control, each lot has been sampled, tested and released by Quality Assurance Department for the following characteristics:

- Burst Pressure Test
- Bubble Point Test
- Pressure Hold Test
- Flow Rate Performance

Kind regards

BRAND GMBH + CO KG

i.A.



Dr. Christian Schurz
Product Manager Glass & Plastic Ware

i.A.



Dr. Nicolas Frank
Product Manager Cell Culture