



Certificate of Conformance

Becton Dickinson and Company
BD Diagnostic Systems
PO Box 999
Sparks, MD 21152-0999 US

Page 1 of 1

Product Name: Cultureswab Plus Amies w/o Charcoal Single Swab

Catalog Number: 220116

Batch Number: 250623600

Expiration Date: 2026/12/22

Manufacture Date: 2025/04/23

Final Release Testing

01 Sterility: This product has been irradiated with a validated dosage range and is in conformity with ISO 11137. Sterility of contents is guaranteed until the expiration date unless product packaging is opened or damaged.

02 Biological Performance: Samples were inoculated with the test organisms and gave cultural responses as indicated:

CULTURE	ATCC®	RECOVERY
Haemophilus influenzae	10211	Good
Neisseria gonorrhoeae	43069	Good
Bacteroides fragilis	25285	Good

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

BD Diagnostic Systems is an ISO 13485:2016 Registered facility. BD Diagnostics - Diagnostic Systems products are manufactured in facilities registered with the United States Food and Drug Administration (FDA), and are regulated by the FDA's Quality System Regulations (QSRs). This product met BD Diagnostics - Diagnostic Systems stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release. This material is not for human or animal consumption.

Fabiano Andreazza
BD Life Sciences - Diagnostic Systems
Senior Manager, Quality Management
Signature Date: 2026/04/29 17:6:56 GMT