



Specification

1.16997.0001 Total Hardness Test Method: reflectometric with test strips 0.1 - 30.0 °d
Reflectoquant®

Specification		
function control	conforms	
Assay (standard 1)	3.1 - 3.9	°dH
Assay (standard 2)	8.8 - 10.8	°dH
Assay (standard 3)	22.7 - 27.7	°dH

Peter Unglaube

Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.