



Bel-Art Products, Inc.
2024 Broad Street Pocomoke City, MD 21851 USA
Tel: 973-694-0500 * Fax: 973-694-7199

CERTIFICATE OF CONFORMANCE

Product #: 36902-0000 **Mfg. Date:** 4/10/2025 **Lot #:** J392594

Description: "Sterileware" Sterile Sampler Spoons

Material: High-Impact Polystyrene Natural, sterilized in individual polyethylene bags.
Materials meet FDA requirements for regulation 21 part 177.1640
White polystyrene Colorant FDA regulations 21 CFR 177.1640, 178.3297, 175.300

Other: Gamma Treatment: 23.5-50.0 kGy
No testing is done for Pyrogens / End toxins at this time.
Based on the process and materials involved in the manufacturing of this product
this product is considered to be BSE/TSE protein free. AAMI guidelines are followed.
Based on the information provided to us this product does not contain or come in
contact with lead, latex, Mercury, Melamine, formaldehyde, or any other contaminants.
To the best of our knowledge based on the information provided to us this product does
not contain or come in contact with any animal origin.
Manufactured: Bel-Art Products Pocomoke City, MD USA

Reviewed and Approved By:

Signature:

Printed Name: Mohamed T Mohamed

Title: Quality Engineer

Date: 5/5/2025