



Healthcare Corporation

MEDICAL PRODUCTS - CLEVELAND FACILITY

911 North Davis Avenue, Cleveland, MS 38732, Phone: 662.843.9421, Fax: 662.846.5987

Certificate of Analysis

Product: STERILE WATER FOR IRRIGATION, USP - 250 ML

Lot #: G115246

Code: 2F7112

Expiry: Nov 2017

Manufactured Date: 11/06/2014

Sterilization Date: 11/06/2014

TEST NUMBER	TEST NAME	LIMIT	RESULT
11-20-15-001	Oxidizable Substances	Pass	Pass
11-21-16-003	pH at 25 deg. C	5.0 - 7.0	5.6
11-22-03-007	Conductivity Determination of Water (USP/EP)	NMT 5 uS/cm	2 uS/cm
20-03-07-006	Particle Analysis - Large Volume (Instrumental)		
	Particles >= 10 micron	NMT 25 particles/mL >=10 microns	3 ct/mL
	Particles >= 25 micron	NMT 3 particles/mL >=25 microns	0 ct/mL
13-01-V	Bacterial Endotoxin	< 0.25 EU/mL = Pass	Pass
14-04-01-120	Sterility	Pass Parametric Release	Pass
PARAMETRIC RELEASE IN ACCORDANCE WITH U.S. FDA REQUIREMENTS			

THIS PRODUCT WAS MANUFACTURED ACCORDING TO CURRENT GMP,
FULLFILLS THE REQUIREMENTS OF THE MASTER PRODUCTION DOCUMENT AND TESTED TO USP COMPLIANCE.

S. Kilchens 11-20-14

Prepared By/Date
SHARON KITCHENS

T. Porter 11-21-14

Quality Management Approval /Date
Twila Porter

