

Certificate of Conformity

VWR Catalogue Number	391-0645
Description	PHP petri plate, stack size 10, with 90 mm Diameter, Sterile
Lot Number	620/0621
Country of Origin	Manufactured in India
Date of Issue (yyyy-mm-dd)	2021-12-20

Quality System Compliance

Products are manufactured under the **ISO 9001:2015** standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

Material	Virgin Crystal Polystyrene; US FDA approved
Dimensional Specification	Base External Diameter 87 ± 1 mm Base Height 13.70 ± 1 mm Lid External Diameter 91.70 ± 1 mm Lid Height 7.52 ± 1 mm Assembled Height 14.65 ± 1 mm Assembled Weight 14.75 ± 1 g
Sterilization	Products are Gamma irradiated as per ISO 11137 which meets a minimum requirement of 10^{-6} SAL with a specified dose range of 10-15 kGy.
ATP Assay	Not Applicable.
DNase & RNase Free	This product is free of any detectable DNase/RNase contamination.

BSE/TSE	No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE).
Cytotoxicity	This product is free of cytotoxicity effects.
Non-Pyrogenic Statement	Not Applicable.
Latex Statement	This product is not made from natural rubber latex.
BPA Statement	Bisphenol are not used in the manufacture of the raw material and are not expected to be present.
DEHP Statement	Not Applicable.
RoHS	Not Applicable.
REACH Statement	Not Applicable.
Storage Conditions	Room temperature.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.