



Product Number: TX3215
Lot Number: 06100425
Manufactured Date: 4/29/2025
Manufacturing Location: USA

Certificate of Compliance

Texwipe certifies that the above product/material listed conforms to procedures and specifications according to Texwipe Operating Procedures and Work Instructions.

Texwipe certifies that the lot number provided under this Certificate has been inspected and tested, in accordance with Texwipe specifications.

THE FOLLOWING INFORMATION IS PRODUCT SPECIFIC

Sterile Products: Texwipe certifies that the lot number provided under this certificate has been gamma irradiated to a sterility assurance level of 10^{-6} according to the guidelines of Association for the Advancement of Medical Instrumentation.

Sterile Alcohol and Sterile Pre-wetted Wipers: Texwipe certifies that the lot number provided under this certificate has been filled with a final concentration as listed in the Certificate of Analysis and USP Purified Water solution filtered through 0.2 micron filters.

Approved By:

Christopher D. Wall
Quality Assurance Manager

Date: September 30, 2025

Time: 1:15:19 PM EST

TEX-QA-REC-8.2-001 Rev001

Confidential

Texwipe Americas
1210 South Park Drive
Kernersville, North Carolina 27284
United States
Tel: +1 336-996-7046
Toll Free: 800-839-9473
E-mail: info@texwipe.com

Texwipe Europe/Middle East/Africa
ITW Contamination Control BV
Saffierlaan 5
2132 VZ Hoofddorp
The Netherlands
Tel: +31-88-1307-410
E-mail: info@itw-cc.com

Texwipe Asia/Australia/Pacific Rim
5, Circuit Street, SEPZ, LISP 1,
Barrio Diezmo, Cabuyao, Laguna,
Philippines, 4025
Tel: +65-9115-8496
E-mail: asia@texwipe.com

Texwipe China
No.4660, Pang Jin Road,
Wujiang Economic & Technological
Development Zone,
Suzhou, Jiangsu, China
Tel: +86-512-6303-3700
E-mail: asia@texwipe.com