

## ISO conformity of Spectroquant® Ammonium Cell Tests

| Article | Measuring range                       |
|---------|---------------------------------------|
| 1.14544 | 0.5 – 16.0 mg/l NH <sub>4</sub> -N    |
| 1.14558 | 0.20 – 8.00 mg/l NH <sub>4</sub> -N   |
| 1.14559 | 4.0 – 80.0 mg/l NH <sub>4</sub> -N    |
| 1.14739 | 0.010 – 2.000 mg/l NH <sub>4</sub> -N |

Hereby we confirm that the above-mentioned Spectroquant® Ammonium Cell Tests fully comply with

**ISO 23695**

**Water quality — Determination of ammonium nitrogen in water — Small-scale sealed tube method**

Merck KGaA, Darmstadt, Germany

Darmstadt, 12.12.2023

This document is computer-generated and is valid without signature.

Merck KGaA, Darmstadt, Germany is ISO 9001, 14004:2015 and ISO 45001, 50001:2018 certified

We provide information and advice to our customers on application technologies and regulatory matters to the best of our knowledge and ability, but without obligation or liability. Existing laws and regulations are to be observed in all cases by our customers. This also applies in respect to any rights of third parties. Our information and advice do not relieve our customers of their own responsibility for checking the suitability of our products for the envisaged purpose.

**© 2023 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.**

Merck, the vibrant M, Supelco, and Spectroquant are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates. All other trademarks are the property of their respective owners. Detailed information on trademarks is available via publicly accessible resources.

The life science business of Merck KGaA, Darmstadt, Germany operates as MilliporeSigma in the U.S. and Canada

Merck KGaA, Frankfurter Straße 250, 64293 Darmstadt, Germany –  
EMD Millipore Corporation, 400 Summit Drive, Burlington MA 01803, USA –  
MilliporeSigma Canada Ltd., 2149 Winston Park Dr, Oakville, Ontario, L6H 6J8, Canada

