

Certificate of Conformity

VWR Catalogue Number	734-2983
Description	VWR® Cell Chambers, 10 levels, 1300 - 2000 ml
Country of Origin	Manufactured in China
Date of Issue (yyyy-mm-dd)	2026-03-02

Quality System Compliance

This certificate confirms that the product has been sourced from a supplier or manufacturer who has declared their Quality Management System (QMS) complies with **ISO 9001:2015 & ISO 13485:2016**.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Test/Procedure	Results
Visual Attributes	Pass
Flatness	Pass
Cell culture	Pass
Packaging	Pass

Product Specifications

Material	Polystyrene (PS), Compliance with USP Class VI Cap - High Density Polyethylene (HDPE), Compliance with USP Class VI Membrane-0.22µm Hydrophobic.
Cell Growth Area	6416 cm ²
Sterilization	Product labelled as sterile is EB (Electron beam) irradiated as per ISO 11137 and meets SAL 10 ⁻⁶ with a specified dose range of 15-30kGy.
ATP Assay	Not Applicable.

DNase & RNase Free	This product is free of any detectable DNase/RNase contamination.
BSE/TSE	No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE) according to EMEA 410/01 revision 3.
Cytotoxicity	Testing is conducted to quality all material resins using ISO 10993 standards for cytotoxicity and have been shown to be non-toxic.
Non-Pyrogenic Statement	The acceptance level for product is 0.06 EU/mL or less than 20 EU/device (TAL Gel Clot Method). Test passed,
Latex Statement	This product is not made from natural rubber latex.
BPA Statement	Bisphenol-A (BPA) and Bisphenol-F (BPF) are not used in the manufacture or the formulation of the raw materials and are not expected to be present. However, tests have not been performed for these chemical materials.
DEHP Statement	DEHP not present.
RoHS	No substances (Lead, Cadmium, Mercury, Hexavalent Chromium (Cr6+), Poly Brominated Biphenyls (PBB), Poly Brominated Diphenyl ethers (PBDE), Bis(2-Ethyhexyl) Phthalate (DEHP), Benzyl Butyl Phthalate (BBP), Dibutyl Phthalate (DBP), Diisobutyl Phthalate (DIBP)) are used in manufacturing the raw materials and final product. No routinely analysis is performed.
REACH Statement	Not Applicable.
Storage Conditions	Room temperature.
Shelf Life	3 years.

Disclaimer: VWR's Quality Management System (QMS) is ISO 9001:2015 certified as a wholesale distributor of finished products, ensuring effective oversight of finished product handling, traceability, and distribution processes. VWR is not a manufacture and does not engage in manufacturing, co-manufacturing or processing activities such as design, packaging, labeling, testing or quality release. Consequently, the supplier or manufacturer's QMS certification may differ from VWR's.

VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. This product is intended for research use only. It is not for use in diagnostic procedures, therapeutic applications, or any human or clinical use.

This document has been produced electronically and is valid without a signature.

Version History

Sl.no	Version number	Description	Month & Year
1	1.0	Created General CoQ using new layout	03-2026