

CORNING GOSSELIN SAS

123 Route de Caëstre

59190 - Borre

FRANCE

www.corning.com/lifesciences

Refer to website for regional contact information

Date of certificate : 2025 - 01 - 14

Page 1

Product description

Container 400mL WhitePP SnapCap Ste rile Assembled 185/CS

Catalogue number

PC400-05

Lot number

901365

Lots : 901365

Quality Management System

Complies with the current version of the ISO9001 standard.

Manufacturing date

2024/08/01

Sterility Assurance Level

Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products - Radiation". Products meet a minimum Sterility Assurance Level (SAL) of 10⁻³.

Expiration date

2027/08/01

Expiration statement

This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release



Sandra Bianchi

Plant Quality Manager, Borre

BianchiS@corning.com