

Certificate of Conformity

VWR Catalogue Number	213-0462
Description	Beakers, low form, heavy-duty, with 50 mL Capacity, 60 mm Height, 42 mm Ext.Diameter
Lot Number	GL1025
Date of Manufacture (yyyy-mm-dd)	2025-03-03
Country of Origin	Manufactured in India
Date of Issue (yyyy-mm-dd)	2025-07-01

Quality System Compliance

Products are manufactured under the [ISO 9001:2015](#) standard and calibrated in [ISO 17025:2017](#) certified Laboratory. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

The product is in conformance with the requirements of the following standards and regulations:

Technical standards	DIN 12331
Calibration procedure	N/A
Material	Borosilicate 3.3 Glass DIN ISO 3585:1998 TYPE-1, CLASS A, and meets ASTM E-438, TYPE-1, CLASS-A Standard.
Sterilization	Hot air sterilization, in the oven up to 140 °C Steam sterilization, in an autoclave 121 °C/ 20 min 134 °C/ 10 min.
ATP Assay	Not applicable.
DNase & RNase Free	Not applicable.

BSE/TSE	No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE).
Cytotoxicity	Not applicable.
Non-Pyrogenic Statement	Not applicable.
Latex Statement	Not applicable.
BPA Statement	Not applicable.
DEHP Statement	Not applicable.
RoHS	Not applicable.
REACH Statement	Not applicable.
Storage Conditions	Stored at room temperature in normal warehouse condition.
Shelf Life	No predefined shelf life.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature