

Certificate Of Processing

Systagenix
Wound Management Ltd.
Gargrave, BD23 3RX, UK
Tel: +44 (0) 1756 747200

The products listed below have been irradiated at Gargrave Sterilisation Services, Systagenix, Gargrave, UK.

Customer: VWR International Ltd

Irradiation Batch Number/Schedule: 40,267

Irradiation Lot Start Date/Time: 09-Jan-24 9:50:36PM

Irradiation Lot End Date/Time: 10-Jan-24 1:49:54AM

Number Of Interruptions: 1

Comment: Product Description: Ergo 500ml Sq

PO: 4515434205

Expiry: 28/12/2025

Inventory Detail

Finished Goods Code	Lot Number	Released Quantity	Delivered Dose (kGy)		Specified Dose Range (kGy)
			Minimum	Maximum	
VWRI331-0280	10653837	60	27.5	34.8	15.0 - 40.0

Gargrave Sterilisation Services/Systagenix certifies that the materials listed above received the absorbed doses within the precision and accuracy of the dosimetry system employed.

Signature Manifest

Signed By Supervisor **Lee Barrett**

Signed On: 10-Jan-24 At: 7:34 AM GMT:0:00

Signed By QA **Sharon Allport**

Signed On: 10-Jan-24 At: 12:39 PM GMT:0:00

Document Content Revision: 1