

**Certificate of Analysis**

**Product: STERILE WATER FOR IRRIGATION, USP - 1000 ML**

**Lot #: G111534**  
**Code: 2F7114**  
**Expiry: FEB 2017**

**Manufactured Date: 02/28/2014**  
**Sterilization Date: 02/28/2014**

| TEST NUMBER                                                 | TEST NAME                                       | LIMIT                            | RESULT  |
|-------------------------------------------------------------|-------------------------------------------------|----------------------------------|---------|
| 11-20-15-001                                                | Oxidizable Substances                           | Pass                             | Pass    |
| 11-21-16-003                                                | pH at 25 deg. C                                 | 5.0 - 7.0                        | 5.6     |
| 11-22-03-007                                                | Conductivity Determination of Water (USP/EP)    | NMT 5 uS/cm                      | 1 uS/cm |
| 20-03-07-006                                                | Particle Analysis - Large Volume (Instrumental) |                                  |         |
|                                                             | Particles >= 10 micron                          | NMT 25 particles/mL >=10 microns | 1 ct/mL |
|                                                             | Particles >= 25 micron                          | NMT 3 particles/mL >=25 microns  | 0 ct/mL |
| 13-01-V                                                     | Bacterial Endotoxin                             | < 0.25 EU/mL = Pass              | Pass    |
| 14-04-01-120                                                | Sterility                                       | Pass Parametric Release          | Pass    |
| PARAMETRIC RELEASE IN ACCORDANCE WITH U.S. FDA REQUIREMENTS |                                                 |                                  |         |

THIS PRODUCT WAS MANUFACTURED ACCORDING TO CURRENT GMP,  
FULLFILLS THE REQUIREMENTS OF THE MASTER PRODUCTION DOCUMENT AND TESTED TO USP COMPLIANCE.

*E. Soliz* 3-4-14  
Prepared By/Date

**ELLEN SOLIZ**

*Twila Proctor* 3-4-14  
Quality Management Approval /Date  
*Twila Proctor*