

Certificate of Conformity

VWR Catalogue Number	See list
Description	VWR®, Plain Microscope Slides
Country of Origin	Manufactured in China
Date of Issue (yyyy-mm-dd)	2025-05-21

Catalogue Number	Designation	Thickness [mm]	L×W [mm]	Pk
631-1160	Ground edges, clipped corner (45°) slides	1.0 (-0.05/+0.1)	76×26	2.500

Quality System Compliance

Products are manufactured under the [ISO 9001:2015](#) standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

Material	Sheet Glass.
Length	75.60~ 75.90 mm
Width	25.00 ~ 25.30 mm
Thickness	1.00 (-0.05/+0.1) mm
Refractive index	1.53 +/-0.02 of refractive index of the light at the green mercury e-line (wavelength=547.07nm) ISO 8037-2.
Wettability	The central parts of the two surfaces of the slides should be wetted equably, and slightly dehydration of the edges is allowed. (ISO 8037-2)

Sterilization	Not applicable.
ATP Assay	Not applicable.
DNase & RNase Free	Not applicable.
BSE/TSE	No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE).
Cytotoxicity	Not applicable.
Non-Pyrogenic Statement	Not applicable.
Latex Statement	Not applicable.
BPA Statement	Not applicable.
DEHP Statement	Not applicable.
RoHS	Not applicable.
REACH Statement	Not applicable.
Storage Conditions	Room temperature.
Shelf Life	5 years.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.

Version History

Sl.no	Version number	Description	Month & Year
1	1.0	Created General CoQ using new layout	05-2025