



Certificate of Conformance, Analysis, and Sterility

This certification is provided as full assurance that the following product code and lot number were manufactured in accordance via a Quality System certified to ISO 9001 with prescribed procedures and specifications.

Product Code:	PSC20001-BPR
Lot Number:	928823-0-0-1
EU BPR Authorisation Number:	EU-0020460 0002 (1-2)
Manufacturing Date:	05/10/2022
Expiration Date:	05/2024

Analysis

% v/v Isopropyl Alcohol:	69.988
--------------------------	--------

Sterility

Gamma Processing Run ID:	3117947
Irradiation Dose Specified:	25 kGy to 50 kGy
Irradiation Dose Delivered:	28.5 kGy to 37.6 kGy
Validated Sterile Product:	Validation Sterile packets available upon request.
Sterility Assurance Level:	10 ⁻⁶ (SAL determined by test method AAMI/ISO 11137)

A handwritten signature in black ink, appearing to read "C. Duffee".

QC Supervisor
coc@contecinc.com

Wednesday, May 18, 2022

NORTH AMERICA

Contec, Inc.
525 Locust Grove
Spartanburg, SC 29303 USA
tel: +1 864 503 8333
fax: +1 864 503 8444

EUROPE

Contec Europe
ZI Du Prat, Avenue Paul Dupleix
56000 VANNES France
tel: +33 (0)2 97 43 76 98
fax: +33 (0) 97 43 76 86

ASIA

Contec Cleanroom Technology (Suzhou) Co., Ltd.
17 Longyun Road
Suzhou Industrial Park, Suzhou 215024 China
tel: +86 512 6274 4050
fax: +86 512 6274 4051

web: www.contecinc.com
e-mail: info@contecinc.com

Page 1 of 1

DID: 20605A



Certificate of Processing

STERIGENICS 10811 Withers Cove Park Dr. Charlotte NC 28278
TEL 704 588-6877 FAX 704 588-8355 www.sterigenics.com

R55480102

05/16/2022 15:21:56 GMT
Page 1 of 1

Customer Name:	Contec- PO Box 530	Processing Facility:	Charlotte Gamma	Work Order #	3117947
P.O.#	125514			Sales Order #	3042807
25-50 kGy		CNT1, Gamma Treatment	Irradiation Date/Time:	05/15/2022 07:10:00 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	192	CA	PSC20001-BPR	Clean Room Product	928823-0-0-1	125514
102.000	108	CA	SB327030IR	32 oz IPA Spray Bottle 70/30	928810-0-0-1	125514
103.000	108	CA	SB327030IR	32 oz IPA Spray Bottle 70/30	928814-0-0-1	125514
104.000	108	CA	SB327030IR	32 oz IPA Spray Bottle 70/30	928815-0-0-1	125514
105.000	72	CA	HCFT7030IR_32	32 oz Sterile IPA w/ cap	928019-0-0-1	125514
106.000	72	CA	HCFT7030IR_32	32 oz Sterile IPA w/ cap	928796-0-0-1	125514
	660	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By ----- User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	50.0 kGy	28.5 KGY	Pass	WKOBLISH WENDY KOBLISH	05/16/2022 13:26:28 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	50.0 kGy	37.6 KGY	Pass	BTOWERY Benjamin Towery	05/15/2022 20:04:41 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: BRITTANY LONG
Reason: Work Order Completions

Date: 05/16/2022 15:20:48 GMT