

Certificate of Conformity

VWR Milk Testing Pipette, Purple Colour Code, Sterile

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.


VWR Catalogue No: 612-5851

Lot No: 201126297

Date of Manufacture: 26th Nov 2020

Expiry Date: 26th Nov 2023

Date of Sterilization: 28th Nov 2020



Material: Pipet: GPPS (General purpose Polystyrene, Compliance with USP Class VI; Colour: Clear).
Filter: PO (Polyolefin)

Volumetric Accuracy: Volumetric Accuracy is calibrated for accurate dispensing to within ±2% of full range.

Quality System Compliance: Products are manufactured under the ISO 9001:2015 & ISO 13485:2016 standard. Products are Inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release. Manufactured in a Class 100 room environment.

Certificate of Conformity

QC Testing: Representative production samples are collected and inspected in accordance with current applicable product specifications. Testing parameters and their results:

Test procedure	Results
Visual Attributes	Pass
Volumetric	Pass
Gastight test	Pass
Ink fastness	Pass
Packaging	Pass

Sterility: Products labeled as sterile is EB (Electron Beam) irradiated and dose released upon ISO11137 recommended practices meets a minimum requirement of 10^{-6} SAL. Products labeled sterile meets a minimum requirement of 10^{-6} SAL with a specified dose range of 15 - 30 kGy.

BSE/TSE Statement: The above product contains no ingredients of animal origin and no material derived from, or exposed to animals affected by or under quarantine for Transmitting, Animal Spongiform Encephalopathy (TSE)/ Bovine Spongiform Encephalopathy (BSE)

Non-pyrogenic: The acceptance level of product is less than 0.5 EU/ml (TAT Gel Clot Method).

DNase & RNase free: This product has been tested and is free of any detectable DNase/RNase contamination.

Cytotoxicity: Testing is conducted to qualify all material resins using ISO 10993 standards for cytotoxicity and have been shown to be non-toxic.

Latex Free statement: Product is Latex Free

Haemolytic statement: These Serological Pipets had no effect on haemolytic properties under this test condition (HR less than 5% considered acceptable).

BPA Statement: Bisphenol-A (BPA) and Bisphenol-F (BPF) are not used in the manufacture or the formulation of the raw materials and are not expected to be present. However, tests have not been performed for these chemical materials.

DEHP Statement: Diethylhexyl phthalate (DEHP) is not included in the manufacture or the formulation of the raw materials and are not expected to be present. However, tests have not been performed for these chemical materials.

Heavy metals - RoHS: Raw materials comply with the requirements of the Directive 2002/95/EC, as amended, and 2011/65/Eu concerning the limits of cadmium, lead, mercury, hexavalent chromium, Polybrominated Biphenyls (PBB) and Polybrominated Diphenyl Ethers (PBDE).

Certificate of Conformity

Food contact statement: The composition of the product complies with the requirements of the 21 CFR 177.1640 "Polystyrene and rubber modified polystyrene".

Storage conditions : Storage at normal temperature.

This document has been produced electronically and is valid without a signature

Date of Issue: **April 2021**