

Certificate of Conformity

VWR Roller Bottle, Treated Surface, Plug seal cap, 2000 ml

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: 734-2750

Lot No: 201222679

Date of Manufacture: 22nd Dec 2020

Expiry Date: 22nd Dec 2023

Date of Sterilization: 24th Dec 2020

Description	Capacity (ml)	Growth area (cm ²)	Packed	Pk
Treated surface, plug seal cap	2000	850	20/bag, 20/case	20

This product is Class I category device as defined by the FDA in 21CFR Parts 862-892

Material: Bottle - Polystyrene (PS)
Cap – High Density Polyethylene (HDPE)

Quality System Compliance: Products are manufactured under the ISO 9001:2015 & ISO13485:2016 standard. Products are Inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

Quality Testing: Representative production samples are collected and inspected in accordance with current applicable product specifications.

Accuracy: Graduation accuracy comply with our product specification $\pm 5\%$.

Certificate of Conformity

Sterility: Products labeled as sterile is EB (Electron Beam) irradiated and dose released upon ISO11137 recommended practices meets a minimum requirement of 10^{-6} SAL with the specified dose range of 15 - 30 kGy.

BSE/TSE Statement: The above product contains no ingredients of animal origin and no material derived from, or exposed to animals affected by or under quarantine for Transmitting, Animal Spongiform Encephalopathy (TSE)/ Bovine Spongiform Encephalopathy (BSE)

Non-pyrogenic: The acceptance level of product is less than 0.5 EU/ml or less than 20 EU/device by TAL Gel Clot Method. Test passed.

DNase & RNase free : This product has been tested and is free of any detectable DNase/RNase contamination.

Cytotoxicity: Testing is conducted to quality all material resins using ISO 10993 standards for cytotoxicity and have been shown to be non-toxic.

Latex Free statement: Product is Latex Free

BPA Statement: Not applicable.

DEHP Statement: Not Applicable.

Heavy metals - RoHS: Raw materials comply with the requirements of the Directive 2002/95/EC, as amended, and 2011/65/Eu concerning the limits of cadmium, lead, mercury, hexavalent chromium, Polybrominated Biphenyls (PBB) and Polybrominated Diphenyl Ethers (PBDE).

This document has been produced electronically and is valid without a signature

Date of Issue: **June 2021**