

Certificate of Quality

VWR Catalogue Number	See List
Description	VWR® Polypropylene Specimen Containers
Country of Origin	Manufactured in USA
Date of Issue (yyyy-mm-dd)	2022-03-24

VWR Catalogue Number	Description	Capacity	Dimensions	Sterility
89107-308	Container	148 mL (5 oz.)	73.7 Dia x 66.3 H mm	Nonsterile
89125-844	Container with Lid	148 mL (5 oz.)	73.7 Dia x 66.3 H mm	Sterile
89107-310	Container	237 mL (8 oz.)	73.7 Dia x 100.5 H mm	Nonsterile
89125-846	Container with Lid	237 mL (8 oz.)	73.7 Dia x 100.5 H mm	Sterile
89107-312	Lid	Nil	Nil	Nonsterile
89125-850	Lid	Nil	Nil	Sterile

Quality System Compliance

Products are manufactured under the [ISO 9001:2015 & ISO 13485:2003](#) standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable product specifications.

Product Specifications

Material **Container:** Polypropylene, 100 % Virgin, FDA 21 CFR Compliance, USP Class VI passed
Lid: Low Density Polyethylene, 100 % Virgin, FDA 21 CFR Compliance, USP Class VI passed

Sterilization Product labelled as sterile is sterilization by B/C Gamma Process with a specified dose range of 20-50 kGy.

ATP Assay Not applicable.

DNase & RNase Free Not applicable.

BSE/TSE	All materials used and contact materials are of fully synthetic origin.
Cytotoxicity	Not applicable.
Non-Pyrogenic Statement	Not applicable.
Latex Statement	Not applicable.
BPA Statement	Not applicable.
DEHP Statement	Not applicable.
RoHS	Not applicable.
REACH Statement	Not applicable.
Storage Conditions	Climate controlled storage.
Shelf Life	3 years.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.