

CERTIFICATE OF CONFORMANCE

Product Part Number: MZCRMS-10IR
Lot Number: 013068-000
Irradiation Reference Number: 0002327-00
Lot Quantity: 480 CASE(S)
Date of manufacture: 04/06/2019
Shelf Life: N/A
Description: MEGASORB MOP, BLUE, IRR'D (12/CS)
Size -
Width: 6.5"
Length: 15"
Height: N/A

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.



04/09/2019

Authorized Signature
Quality Control Department

Date

CC: With Product
Customer File
C of C File

Certificate Of Processing

Prepared for MICRONOVA MFG INC



Gamma Process Run ID 114587A

<u>Product Code</u>	<u>Product Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
MZCRMS-10IR	013068-000	480	CS

Other Information:

0002327-00

Processing Run Start Date/Time:	05-Apr-2019 03:30:09 pm	Approx. Downtime (hours):	1.13
Processing Run End Date/Time:	06-Apr-2019 12:15:50 am		

Minimum Specified Dose (kGy):	40.0	Minimum Delivered Dose (kGy):	44.1
Maximum Specified Dose (kGy):	60.0	Maximum Delivered Dose (kGy):	55.3

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Signature Manifest



Reviewed and E-Signed By

Uprety, Amit (QS & RC Analyst I)

Document Content Revision: 1

Signed On 4/6/2019 at 1:59 PM

UTC / GMT Offset (hh:mm): -7:00

Processing Location:

STERIS
1000 South Sarah Place
Ontario, CA 91761
Phone: 909-390-9942
Fax: 909-390-7854

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485, and in alignment with EN ANSI/AAMI/ISO 11137. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.