

Certificate of Quality



Part Number	MFLX95644-00
Lot Number	28368155
Part Description	SAMPLING SYSTEM 125ML
Expiration Date	03MAY2025
Country of Origin	USA

Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Product Specifications & Compliance

- Compound: N/A
- The C-Flex tubing used in this assembly meets the requirements of USP, Biological Test for Plastics, and Class VI.
- This assembly has been gamma irradiated within the range of 25.0 kGy to 40.0 kGy.
- This product is not approved or intended for, and should not be used for medical, clinical, surgical or other patient oriented applications.
- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.