



Certificate of Conformity

VWR Pipette Tip, Reload System 100-1250 µL

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: 613-1062

Lot No: 20150298

PCR inhibitors: not present

Deoxyribonucleic Acid (DNA): < 30pg

Ribonuclease (RNase): < 10⁻⁹ Kunitz Units/µl

Deoxyribonuclease (DNase): < 10⁻⁷ Kunitz Units/µl

Endotoxin (Pyrogen): Product is certified non-pyrogenic with an endotoxin level ≤ 0.06 EU/ml

Adenosine-5'-triphosphate (ATP): < 10⁻¹³ mg/µl

This document has been produced electronically and is valid without a signature

Date of issue: **November 2020**