



OKC Allergy Supplies, Inc.

(formerly Allergy Laboratories, Inc.)

a division of

ALK-Abello, Inc.

Certificate of Analysis

Product:	SEV – Sterile Empty Vial
Lot number:	SEV2012919
Vial size:	30mL-20mm
Manufacture date:	01/29/2019
Part #:	SEV30

Quality Release Testing		
Assay	Specifications	Results
Appearance-product	Essentially free of visible particulate	Pass
Appearance-packaging	No apparent leakage or physical alteration	Pass
Sterility, USP <71>	Sterile (no microbial growth)	Conforms
Bacterial Endotoxins, USP <85>	≤ 0.25 EU/mL	< 0.005 EU/mL

Release Date	Expiration Date
03/06/2019	01-2023

Quality Assurance Approval	
By / Date:	<i>Liberty E. Eneanya</i> 04-17-19

Original Certificate

1005 S.W. Second Street
Oklahoma City, OK 73109

Tel: (800) 654-3971
Fax: (800) 811-3389

www.allergylabs.com
KS030719