

CERTIFICATE OF CONFORMANCE AND IRRADIATION

This certification is provided as full assurance that the following product code was manufactured in accordance with prescribed procedures and specifications.

Product Code: MHQT0150

Lot Number: 015610

Manufacturing Date: 2020.07.27

Expiration Date: 2022.07

Radiation Date: 2020.08.05

Radiation Lot: 200805204

Irradiation Dose Specified: 25 kGy to 50 kGy

Minimum Absorbed Dose: 26.9 kGy

Maximum Absorbed Dose: 37.4 kGy

Quality Department
Contec Cleanroom Technology (Suzhou) Co., Ltd.
2020-08-12



苏州中核华东辐照有限公司
Suzhou CNNC Huadong Radiation Co., Ltd.

文件编号 (Doc No.): CHF/QR-8.2.6-02

辐照证明书
CERTIFICATE OF IRRADIATION

执照登记号 (License No.): 913205097186218342 Serial No.: SZKT200051

客户名称 (Customer): 康泰无尘科技 (苏州) 有限公司

辐照日期 (Radiation Date): 2020/08/05 辐照加工批号 (Radiation Lot): 200805204

产品名称 Description	订单号 Order No.	产品货号 (代号) Stock No. (Code No.)	产品批号 Lot No.	箱数 Cases
涤纶针织无菌拖把头套	N/A	MHQT0150	015610	200
棉签	N/A	SF-56IR	015619	1500
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A

客户要求剂量 (Sterilization Dose Required by Customer):

最低剂量 (Minimum Dose): 25.0 kGy 最高剂量 (Maximum Dose): 50.0 kGy

苏州中核华东辐照有限公司特此证明上述产品已经钴-60射线辐照。
Suzhou CNNC Huadong Radiation Co., Ltd. hereby certifies that above products have been irradiated by cobalt-60 gamma ray.

剂量计名称 (Dosimeter Name): 重铬酸钾 (银) 剂量计批号 (Dosimeter Lot): 200221

剂量测量不确定度 (Dosimetry Uncertainty): 0.836%

产品实际最低吸收剂量 (Minimum absorbed dose): 26.9 kGy

产品实际最高吸收剂量 (Maximum absorbed dose): 37.4 kGy

签发 (Issued by): [Signature] 日期 (Date): 2020.8.6

复核 (Reviewed by): [Signature] 日期 (Date): 2020.8.6

本公司通过了ISO13485体系和ISO11137标准的认证。Our company is certified for conformance to ISO13485 and ISO11137 standards.
地址: 江苏省苏州市吴江经济开发区交通路4756号 Address: 4756 Jiaotong Road, Economic and Technological Development Zone of Wujiang, Suzhou, Jiangsu Province, PRC