



CERTIFICATE OF CONFORMANCE AND STERILIZATION

Date: 27-Sep-19

This certification is provided as full assurance that the following product code was manufactured in accordance with prescribed procedures and specifications.

Product/Material Number:	C2-1212IR
Product Description:	STERILESORB* C/P 12INx12IN
Lot Number:	171814
Irradiation Run Number:	2517475
Irradiation Dose Delivered:	33.4kGy to 43.3kGy
Manufacture Date:	09/2019
Expiration Date:	09/2021

The above product has been Gamma Irradiated to Sterility Assurance Level of 10^{-6} in accordance with the guidelines of the Association for the Advancement of Medical Instrumentation.

A handwritten signature in black ink, appearing to read "C. Inthavongsa".

Champagne Inthavongsa
Quality Supervisor
cinthavongsa@contecinc.com

R55480102

 09/23/2019 13:26:31 GMT
 Page 1 of 1

Customer Name:	FG Clean Wipes- Chicopee, MA.	Processing Facility:	Charlotte Gamma	Work Order #	2517475
P.O.#	37907			Sales Order #	2325240
	25-50 kGy	LYM2, Gamma Treatment	Irradiation Date/Time:	09/20/2019 18:07:00 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	200	CA	C2-1212IR	Wiping Cloth	171814	NA
102.000	840	CA	C2-991R/25	Wiping Cloth	171808	NA
	1040	CA	Total			

Processing Summary

Op#	Operation Name	Location	Date/Time In (GMT)	Date/Time Out (GMT)
400.00	Receive Product	UNPROC	09/16/2019 16:47:00	09/16/2019 19:27:57
410.00	Prepare Product	UNPROC	09/16/2019 19:27:57	09/20/2019 01:36:45
420.00	Load Product - A123	UNPROC	09/20/2019 01:36:45	09/20/2019 18:07:00
430.00	Process Product	UNPROC	09/20/2019 18:07:00	09/20/2019 21:36:45
440.00	Unload Product	UNPROC	09/20/2019 21:36:45	09/21/2019 05:41:00
450.00	Collect Run Data	UNPROC	09/21/2019 05:41:00	09/21/2019 09:46:32

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	50.0 kGy	33.4 KGY	Pass	BTOWERY	09/21/2019 09:45:35 GMT
		Reason Code Test				Benjamin Towery	
450.00	Maximum Dose	25.0 kGy	50.0 kGy	43.3 KGY	Pass	BTOWERY	09/21/2019 09:45:53 GMT
		Reason Code Test				Benjamin Towery	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: WENDY KOBLISH
 Reason: Work Order Completions

Date: 09/23/2019 13:25:34 GMT