

Certificate of Conformity

VWR Catalogue Number	634-0675
Description	Disposable cuvettes with 10 mm Path Length, 340nm - 900nm Range, 2.5mL – 4.5mL Volume
Lot Number	0000002887
Country of Origin	Manufactured in Germany
Date of Issue (yyyy-mm-dd)	2022-02-08

Quality System Compliance

Products are manufactured under the **ISO 9001:2015** standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

Material	Polystyrene (PS). The external dimension of disposable cuvettes macro are manufactured according to DIN 58963 part 3. Polystyrene Cuvettes: Standard deviation at 360nm $\leq \pm 0.005$ extinction units.
Sterilization	Non-Sterile.
ATP Assay	Not Applicable.
DNase & RNase Free	This product is free of any detectable DNase/RNase contamination.
BSE/TSE	A small quantity of a substance from animal origin (tallow derived substance) was introduced during the manufacturing process.
Cytotoxicity	Not Applicable.

Non-Pyrogenic Statement	Not Applicable.
Latex Statement	This product is not made from natural rubber latex.
BPA Statement	Bisphenol are not used in the manufacture of the raw material and are not expected to be present.
DEHP Statement	Please see RoHS information below.
RoHS	No substances (Lead, Cadmium, Mercury, Hexavalent Chromium (Cr6+), Poly Brominated Biphenyls (PBB), Poly Brominated Diphenyl ethers (PBDE), Bis(2-Ethyhexyl) Phthalate (DEHP), Benzyl Butyl Phthalate (BBP), Dibutyl Phthalate (DBP), Diisobutyl Phthalate (DIBP)) are used in manufacturing the raw materials and final product. No routinely analyse is performed.
REACH Statement	No substances as listed on >Substance of Very High Concern< candidate list published 2021-07-08 are used in manufacturing of the raw materials used to produce this product. No routinely analyse is performed.
Storage Conditions	Should be stored using the original packaging, at temperature between 10°C and 50°C. Relative Humidity between 30% and 65%.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.

Annex

Filling Volume:

Macro Cuvette min 2.5 mL: max 4.5 mL

Dimensions:

Cuvette 12.5 x 12.5 x 45 mm

Window:

Macro Cuvette 10 x 35 mm

Light path 10 mm