



415 S Airpark Road
Cottonwood, AZ 86326, USA
Phone: (800) 733-0266
Fax: (928) 649-2306

Email: info@preclaboratories.com

CERTIFICATION OF CONFORMANCE

Product Code: 150
Product Description: Cobalt Chloride Test Papers
Lot Number: 082024
Expiration Date: 09/2026

Prior to release for sale all lots of Cobalt Chloride Test Papers are tested in accordance to established quality standards. Test results must be within specifications for the lot to be released.

Each test paper is sensitive to relative humidity conditions above 50%. The papers will turn pink in the presence of water vapor. The higher the relative humidity present, the faster the test papers react.

When stored properly (in a cool, dry place out of direct sunlight) the test papers are stable for a minimum of 2 years from the date of purchase.

This lot was tested and found to meet specifications. It is released for sale.

Kami Roti
President
Precision Laboratories, Inc.

Date: 02/26/26