Certificate of Conformity

VWR Catalogue Number	631-1551
Description	Microscope slides VWR®, Clear, white Glass, Slide with cut edges and frosted end, 1.0 mm thickness, 76 mm length, 26 mm width
Lot Number	031822
Date of Manufacture (yyyy-mm-dd)	2022-03-18
Country of Origin	Manufactured in China
Date of Issue (yyyy-mm-dd)	2022-08-15

Quality System Compliance

Products are manufactured under the ISO 9001:2015 standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

Material	Float Glass.
Refractive index	1.53 +/-0.02 of refractive index of the light at the green mercury e-line (wavelength=547.07nm) ISO 8037-2.
Wettability	The central parts of the two surfaces of the slides should be wetted equally, and slightly dehydration of the edges is allowed. (ISO 8037-2)
Sterilization	Not applicable.
ATP Assay	Not applicable.
DNase & RNase Free	Not applicable.

BSE/TSE	No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE).
Cytotoxicity	Not applicable.
Non-Pyrogenic Statement	Not applicable.
Latex Statement	Not applicable.
BPA Statement	Not applicable.
DEHP Statement	Not applicable.
RoHS	Not applicable.
REACH Statement	Not applicable.
Storage Conditions	Room temperature, dry.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.