

Certificate of Conformity

VWR Catalogue Number	662-1805
Description	Speciality pH electrodes, with 1 m fixed cable for Agar, microbiological samples, cheese, butter, water, pH range: 0 to 14, Flat extra wide, double junction (HDPE/Pellon) type, +10...+100 °C temperature, Electrolyte Gel, 15 mm diameter, 115 mm length, BNC connection
Lot Number	3420
Country of Origin	Manufactured in USA
Date of Issue (yyyy-mm-dd)	2022-03-10

Quality System Compliance

Products are manufactured under the **ISO 9001:2015** standards. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative product sensors are individually tested according to quality procedure DOP-Q126.

Product Specifications

Material	Epoxy.
Sterilization	Product is chemical treated and they meet a minimum requirement of 10 ⁻⁶ Sterility assurance level.
ATP Assay	Not Applicable.
DNase & RNase Free	Not Applicable.
BSE/TSE	Not Applicable.
Cytotoxicity	Not Applicable.
Non-Pyrogenic Statement	Not Applicable.
Latex Statement	Not Applicable.

BPA Statement	Not Applicable.
DEHP Statement	Please see RoHS information below.
RoHS	No substances (Lead, Cadmium, Mercury, Hexavalent Chromium (Cr6+), Poly Brominated Biphenyls (PBB), Poly Brominated Diphenyl ethers (PBDE), Bis(2-Ethyhexyl) Phthalate (DEHP), Benzyl Butyl Phthalate (BBP), Dibutyl Phthalate (DBP), Diisobutyl Phthalate (DIBP)) are used in manufacturing the raw materials and final product. No routinely analyse is performed.
REACH Statement	No substances as listed on >Substance of Very High Concern< candidate list published 2021-07-08 are used in manufacturing of the raw materials used to produce this product. No routinely analysis is performed.
Storage Conditions	0° C - 50° C, wet.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.