

CERTIFICATE OF ANALYSIS**ELIMINase®**

Lot # 0J025

Date of Manufacture October 2020

Recommended Retest date is two years from the date of manufacture

SIZE	250 ml (8 oz.)	950ml (32 oz.)
DECON #	1101	1102

Characteristic	Specifications	Result
Appearance	Clear to straw	Pass
pH	> 11.0	12.4
SP. GR.	Report Value	1.000

Reviewed and Approved by:

Date: October 2020

Karen Wend
Manager, Quality and Technical Services**BSE/TSE Statement:** There are no materials of animal origin in the raw materials or in the process aides used to manufacture any Decon products.