Certificate of Conformity

VWR Catalogue Number	232-0219
Description	Dressing forceps, Purple Colour, with 110 mm Length, 1.5 mm Thickness, 9.9 mm Width, Sterile
Lot Number	20220301
Date of Manufacture (yyyy-mm-dd)	2022-02-27
Expiry Date (yyyy-mm-dd)	2025-03-01
Country of Origin	Manufactured in China
Date of Issue (yyyy-mm-dd)	2024-02-05

Quality System Compliance

Products are manufactured under the **ISO 13485:2012** standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Inspected according to AQL level S-2 1.5 sampling plan per ISO 2859-1:1999.

Product Specifications

The product is Class I category device as defined by FDA in 21CFR Parts 862-892.

Material Acrylonitrile Butadiene Styrene.

Sterilization Product labeled as sterile is Ethylene Oxide (EO) irradiated according to AAMI/ISO 11135-1:2007 recommended practices in effect at the time of validation. Products labeled sterile meets a minimum requirement of 10^{-6} SAL.

ATP Assay Not Applicable.

DNase & RNase Free Not Applicable.

BSE/TSE	No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE).
Cytotoxicity	Not Applicable.
Non-Pyrogenic Statement	Not Applicable.
Latex Statement	Not Applicable.
BPA Statement	Not Applicable.
DEHP Statement	Not Applicable.
RoHS	Not Applicable.
REACH Statement	Not Applicable.
Storage Conditions	Store at room temperature**.**

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.