

Certificate of Analysis

Product: STERILE WATER FOR IRRIGATION, USP - 1000 ML

Lot #: G109991

Code: 2F7114

Expiry: NOV 2016

Manufactured Date: 11/02/2013

Sterilization Date: 11/02/2013

TEST NUMBER	TEST NAME	LIMIT	RESULT
11-20-15-001	Oxidizable Substances	Pass	Pass
11-21-16-003	pH at 25 deg. C	5.0 - 7.0	5.8
11-22-03-007	Conductivity Determination of Water (USP/EP)	NMT 5 uS/cm	1 uS/cm
20-03-07-006	Particle Analysis - Large Volume (Instrumental)		
	Particles >= 10 micron	NMT 25 particles/mL >=10 microns	1 ct/mL
	Particles >= 25 micron	NMT 3 particles/mL >=25 microns	0 ct/mL
13-01-V	Bacterial Endotoxin	< 0.25 EU/mL = Pass	Pass
14-04-01-120	Sterility	Pass Parametric Release	Pass
PARAMETRIC RELEASE IN ACCORDANCE WITH U.S. FDA REQUIREMENTS			

THIS PRODUCT WAS MANUFACTURED ACCORDING TO CURRENT GMP,
FULLFILLS THE REQUIREMENTS OF THE MASTER PRODUCTION DOCUMENT AND TESTED TO USP COMPLIANCE.

E. Soliz 11.5.13

Prepared By/Date

ELLEN SOLIZ

T. Procter 11-6-13

Quality Management Approval /Date

T. Procter