

Certificate Of Processing

Systagenix
Wound Management Ltd.
Gargrave, BD23 3RX, UK
Tel: +44 (0) 1756 747200

The products listed below have been irradiated at Gargrave Sterilisation Services, Systagenix, Gargrave, UK.

Customer: VWR International Ltd

Irradiation Batch Number/Schedule: 29,729

Irradiation Lot Start Date/Time: 22-Sep-21 9:04:20AM

Irradiation Lot End Date/Time: 22-Sep-21 12:46:10PM

Number Of Interruptions: 0

Comment: Product Description: Ergo 500ml Sq

PO: 4513883007/504

Expiry: 28/08/2023

Inventory Detail

Finished Goods Code	Lot Number	Released Quantity	Delivered Dose (kGy)		Specified Dose Range (kGy)
			Minimum	Maximum	
VWRI331-0268	500A/77	60	22.6	35.8	15.0 - 40.0

Gargrave Sterilisation Services/Systagenix certifies that the materials listed above received the absorbed doses within the precision and accuracy of the dosimetry system employed.

Signature Manifest

Signed By Supervisor **Lee Barrett**

Signed On: 23-Sep-21 At: 7:48 AM GMT:1:00

Signed By QA **Sharon Allport**

Signed On: 23-Sep-21 At: 9:57 AM GMT:1:00

Document Content Revision: 1